



SURAT PERINTAH KERJA

INSTALASI PEMELIHARAAN SARANA PRASARANA

NOMOR :

1312/5/26/E/IPSR

KEPADA :

- | | | |
|----------------|---------|------------|
| 1. Pak Khoirul | 3. Aris | 5. Jesisca |
| 2. Pak Syaiful | 4. Dika | |

Selesai IPSRS

TANGGAL LAPORAN	NOMOR LAPORAN	PELAKSANAAN	LOKASI
2026-05-20 13:37:07	188/5/26	Jum'at, 22 Mei 2026	Cendrawasih

NAMA ALAT	MERK	MODEL	NOMOR SERI
Regulator O2		-	

LAPORAN	TINDAKAN	KETERANGAN
regulator oksigen transport ngobos	perbaikan seal karet	



SPAREPART

- | | | |
|--------------|-------|----|
| 1. flowmeter | 1 set | 4. |
| 2. | | 5. |
| 3. | | 6. |

Mengetahui,

Teknisi Pelaksana

Bu Nova

Aris

Dika

Jesisca

Penerima Pekerjaan

Bu Erina

bag. IPSRS RSUD SLG

Admin

Alivia Ayu Ninditya

Ka. IPSRS

Nova Dwi Hapsari
NIP. 198403012011012016