



SURAT PERINTAH KERJA

INSTALASI PEMELIHARAAN SARANA PRASARANA

NOMOR :

1367/6/26/E/IPSR

KEPADA :

1. Pak Khoirul
2. Pak Syaiful
3. Jesisca
4. Aris
5. Dika

Selesai IPSRS

TANGGAL LAPORAN	NOMOR LAPORAN	PELAKSANAAN	LOKASI
2026-05-26 07:49:38	242/5/26	Selasa, 2 Juni 2026	RR

NAMA ALAT	MERK	MODEL	NOMOR SERI
Regulator O2	-	-	-

LAPORAN	TINDAKAN	KETERANGAN
Breathing circuit robek 1 biji, humidifier O2 transport bocor (di rr)	Memberikan breathing circuit baru 1 buah dan mengganti chamber regulator yang bocor	-



SPAREPART

- 1.
- 2.
- 3.
- 4.
- 5.
- 6.

Mengetahui,

Teknisi Pelaksana

Bu Nova

Jesisca

Aris

Dika

Penerima Pekerjaan

Nia

bag. IPSRS RSUD SLG

Admin

Alivia Ayu Ninditya

Ka. IPSRS

Nova Dwi Hapsari
NIP. 198403012011012016