



SURAT PERINTAH KERJA

INSTALASI PEMELIHARAAN SARANA PRASARANA

NOMOR :

1371/6/26/E/IPSRs

KEPADA :

- | | | |
|----------------|------------|---------|
| 1. Pak Khoirul | 3. Jesisca | 5. Dika |
| 2. Pak Syaiful | 4. Aris | |

Selesai IPSRS

TANGGAL LAPORAN	NOMOR LAPORAN	PELAKSANAAN	LOKASI
2026-05-29 07:08:15	258/5/26	Selasa, 2 Juni 2026	Lab Patologi Anatomi

NAMA ALAT	MERK	MODEL	NOMOR SERI
semi enclised Tissue processor	Lei	-	8533

LAPORAN	TINDAKAN	KETERANGAN
ups bunyi	Alarm BT kondisi baterai rusak digantikan baterai sudah normal	sudah normal



SPAREPART

- | | |
|----|----|
| 1. | 4. |
| 2. | 5. |
| 3. | 6. |

Mengetahui,

Teknisi Pelaksana

Penerima Pekerjaan

Bu Nova

Jesisca

Aris

Dika

bag. IPSRS RSUD SLG

Admin

Ka. IPSRS

Alivia Ayu Ninditya

Nova Dwi Hapsari
NIP. 198403012011012016