



SURAT PERINTAH KERJA

INSTALASI PEMELIHARAAN SARANA PRASARANA

NOMOR :

1408/6/26/E/IPSR

KEPADA :

- | | | |
|----------------|---------|-------------|
| 1. Pak Khoirul | 3. Aris | 5. Jেসিস্কা |
| 2. Pak Syaiful | 4. Dika | |

Selesai IPSRS

TANGGAL LAPORAN	NOMOR LAPORAN	PELAKSANAAN	LOKASI
2026-06-03 09:42:05	24/6/26	Kamis, 4 Juni 2026	CSSD

NAMA ALAT	MERK	MODEL	NOMOR SERI
Autoclave	ajc	-	608

LAPORAN	TINDAKAN	KETERANGAN
saluran air bocor di dalam mesin autoclave ajc	kencangkan konektor saluran air	



SPAREPART

- | | |
|----|----|
| 1. | 4. |
| 2. | 5. |
| 3. | 6. |

Mengetahui,

Teknisi Pelaksana

Bu Nova

Aris

Dika

Jেসিস্কা

Penerima Pekerjaan

Aning

bag. IPSRS RSUD SLG

Admin

Alivia Ayu Ninditya

Ka. IPSRS

Nova Dwi Hapsari
NIP. 198403012011012016