



SURAT PERINTAH KERJA

INSTALASI PEMELIHARAAN SARANA PRASARANA

NOMOR :

246/1/26/E/IPSR

KEPADA :

1. Pak Khoirul
2. Pak Syaiful
3. Aris
4. Dika
5. Jesisca

Selesai IPSRS

TANGGAL LAPORAN	NOMOR LAPORAN	PELAKSANAAN	LOKASI
2026-01-26 15:06:46	283/1/26	Selasa, 27 Januari 2026	IBS

NAMA ALAT	MERK	MODEL	NOMOR SERI
O2 Sentral	sme	-	*

LAPORAN	TINDAKAN	KETERANGAN
Oksigen Central Ok 1 bocor	ganti selang yang bocor	



SPAREPART

1. Selang PU 5x8 mm 30 cm
- 2.
- 3.
- 4.
- 5.
- 6.

Mengetahui,

Teknisi Pelaksana

Bu Nova

Aris

Dika

Jesisca

Penerima Pekerjaan

Ardi

bag. IPSRS RSUD SLG

Admin

Alivia Ayu Ninditya

Ka. IPSRS

Nova Dwi Hapsari
NIP. 198403012011012016