



SURAT PERINTAH KERJA

INSTALASI PEMELIHARAAN SARANA PRASARANA

NOMOR :

381/2/26/E/IPRS

KEPADA :

1. Pak Khoirul
2. Pak Syaiful
3. Dika
4. Jesisca
5. Aris

Selesai IPSRS

TANGGAL LAPORAN	NOMOR LAPORAN	PELAKSANAAN	LOKASI
2026-01-28 19:16:53	306/1/26	Minggu, 8 Februari 2026	Kasuari

NAMA ALAT	MERK	MODEL	NOMOR SERI
Bedhead	-	-	-

LAPORAN	TINDAKAN	KETERANGAN
outlet oksigen kamar 10 bed 1 macet tidak bisa dipakai	pengecekan outlet oksigen	pengecekan outlet oksigen alat normal



SPAREPART

- 1.
- 2.
- 3.
- 4.
- 5.
- 6.

Mengetahui,

Teknisi Pelaksana

Bu Nova

Dika

Jesisca

Aris

Penerima Pekerjaan

Arisa

bag. IPSRS RSUD SLG

Admin

Alivia Ayu Ninditya

Ka. IPSRS

Nova Dwi Hapsari
NIP. 198403012011012016