



SURAT PERINTAH KERJA

INSTALASI PEMELIHARAAN SARANA PRASARANA

NOMOR :

77/1/26/E/IPSRS

KEPADA :

1. Pak Khoirul
2. Pak Syaiful
3. Dika
4. Aris
5. Jেসিসca

Selesai IPSRS

TANGGAL LAPORAN	NOMOR LAPORAN	PELAKSANAAN	LOKASI
2026-01-05 08:56:58	36/1/26	Kamis, 8 Januari 2026	Klinik Gigi Prostodonsia

NAMA ALAT	MERK	MODEL	NOMOR SERI
Dental Unit	ANDINI	-	1804012047

LAPORAN	TINDAKAN	KETERANGAN
3 way syringe sudah mulai seret	alat dilakukan pengecekan : pelumasan saklar on of water & air three way syringe	alat dilakukan pengecekan : pelumasan saklar on of water & air three way syringe

rekom : pengeringan setelah digunakan, agar tidak berkerak



SPAREPART

- 1.
- 2.
- 3.
- 4.
- 5.
- 6.

Mengetahui,

Teknisi Pelaksana

Bu Nova

Dika

Aris

Jেসিসca

Penerima Pekerjaan

BU AFIN

bag. IPSRS RSUD SLG

Admin

Alivia Ayu Ninditya

Ka. IPSRS

Nova Dwi Hapsari
NIP. 198403012011012016